

本實驗室在 2015/5/1 依據 AAMI 修改內毒素檢測項目

異動項目如下:

1. Endotoxin 最高容許濃度由 2 EU/ml 改為 0.25 EU/ml

2. 檢測規範如下表

Table 3 is adapted from ISO 13959:2009.

Table 3 — Maximum allowable levels for total viable microbial count (TVC) and endotoxins in dialysis water^a

Contaminant	Maximum allowable level	Action level ^b
TVC	<100 CFU/ml	50 CFU/ml
Endotoxin	<0.25 EU/ml	0.125 EU/ml

^a The reader is cautioned to refer to the latest version of ISO 13959 to ensure that there have been no changes to the values presented in this table.

^b Typically set at 50 % of the maximum allowable level. Lower values may be set.